

Quotation Notice No. 2/23-24 dt. 25-8-23

Sealed Quotations are invited from the manufacturers/authorised distributors for the supply of **Biosafety Cabinet** for implementing the activities envisaged under Plan schemes for the year 23-24.

Size of table SSTOP	Outlet dimension	General requirement
2x2x2 feet	37x32x76 inches	Recirculation of air through a second HEPA Filter. Air drawn through prefilter. HEPA filter having efficiency of 0.3 μ particles with balanced motor and blower assembly (1/5HP, 1440rpm). Air outlet exhaust duct with motor (1/4HP, 2880rpm). 1 UV light for work area.

Biosafety Cabinet has to be delivered and installed at IRTC Lab within seven days from the receipt of supply order. The bills will be settled only after successful supply and installation of equipment.

Last date of submission of offer 14-9-23, 5PM.

